



# Qualikems Lifesciences Pvt. Ltd.

Ho & Works : Plot No. 68-69,G.I.D.C Industrial Estate, Nandesari, Vadodara-391340 (Guj.)

Telefax : 91-265-2841531,2841532,2841534,2841535

Email: salesindia@qualikems.com ,www.qualikems.com

## CERTIFICATE OF ANALYSIS

<b>Name of Product</b>	: Buffer Capsule pH7.0	<b>CAS No.</b>	: --
<b>Molecular Formula</b>	: --	<b>Molecular Weight</b>	: --
<b>Batch No.</b>	: QB2605011	<b>Grade</b>	: LR
<b>Mfg. Date</b>	: May-2026	<b>Retest Date</b>	: April -2031

Sr. No.	Tests	Specification	Observation
1	One capsule dissolved in 100 ml of distilled water	Solution Having pH 7.0±0.05	7.04

**Conclusion:** Product complies as per above specification.

**Note:** This document has been produced electronically and it is valid without signature.

Quality of our product tested and reported in the COA are as per the specification and without any prejudice/liability what so ever the end product performance.

Sales Office : 5531,Basti Harphool singh Sadar Thana Road, Delhi-110006, INDIA

Tel : +91-11-23618475/23618476 | Fax:+91-11-23678476



fssai

